



CooperVision Contact lens parameters

CooperVision Daily disposable contact lenses

| Contact Lens Name | Contact Lens Specifications | Contact Lens powers | Astigmatism powers and axis available |
|--|---|--|--|
| Myday | Base Curve: 8.4 Diameter: 14.2 Dk/t @ 3.00D 100 | +0.25 to +5.00 (0.25 steps) +5.00 to +6.00 (0.50 steps) -0.25 to -6.00 (0.25 steps) -6.00 to -10.00 (0.50 steps) | |
| Proclear 1-day | Base Curve: 8.7 Diameter: 14.2 Dk/t @3.00D 36.6 | +0.25 to +5.00 (0.25 steps) +5.00 to +8.00 (0.50 steps) -0.25 to -6.00 (0.25 steps) -6.00 to -12.00 (0.50 steps) | |
| Proclear 1-day Multifocal | Base Curve: 8.7 Diameter: 14.2 Dk/t @3.00D 28 | +0.25 to +6.00D (0.25 steps) -0.25 to -6.00D (0.25 steps) -6.00 to -10.00D (0.50 steps) Single power profile up to +2.50 | |
| Biomedics 1 day Extra | Base Curve: 8.6 (-ve), 8.8 (+ve) Diameter: 14.2 Dk/t @3.00D 27 | +0.25 to +5.00 (0.25 steps) +5.00 to +6.00D(0.50steps) -0.25 to -6.00D (0.25 steps) -6.00 to -10.00 (0.50 steps) | |
| Biomedics 1 day Extra toric | Base Curve: 8.7 Diameter: 14.5 Dk/t @3.00D 17 | Plano to -6.00D(0.25 steps) -6.50 to -10.00(0.50 steps) | CYL : -0.75, -1.25, -1.75 AXIS : 20°, 90° 160°, 180° (-7.50 to -10.00 AXIS : 90 & 180 ONLY) |



<http://optometryportal.com/>
Follow on Twitter @optometryportal
Download the 'Optometry portal' Android App



Coopervision Contact lens parameters

Coopervision Monthly contact lenses 1/2

| Contact Lens Name | Contact Lens Specifications | Contact Lens powers | Astigmatism powers and axis available |
|----------------------------------|---|--|---|
| Proclear Sphere (monthly) | Base Curve: 8.6 & 8.2 Diameter: 14.2 Dk/t @-3.00D 52.6 | For BC 8.6: +20.00D to -20.00D (0.50D steps after +/-6.00D) For BC 8.2: plano to -10.00D (0.50D steps after +/-6.00D) | |
| Proclear Toric | Base Curve: 8.8 & 8.4 Diameter: 14.4 Dk/t @-3.00D 30.9 | +0.25D to +6.00D (0.25D steps) plano to -6.50D (0.25D steps) -6.50D to -8.00D (0.50D steps) | Cyl: -0.75 -1.25, -1.75, -2.25 Axis: Full Circle (10 deg steps) |
| Proclear Toric XR | Base Curve: 8.8 & 8.4 Diameter: 14.4 Dk/t @-3.00D 30.9 | +6.25 to +10.00 (0.50 steps after +6.50) -8.50D to -10.00 (0.50 steps) | Cyl: -0.75 to -5.75 (-0.50 DC steps) Axis : Full Circle (5 Deg steps) |
| Proclear Multifocal | Base Curve: 8.7 Diameter: 14.4 Dk/t @-3.00D 17 Centre Distance & Centre Near | +0.25D to +6.00D (0.25D steps) plano to -6.50D (0.25D steps) -6.50D to -8.00D (0.50D steps) ADD powers +1.00, +1.50,+2.00,+2.50 | |



<http://optometryportal.com/>
Follow on Twitter @optometryportal
Download the 'Optometry portal' Android App



Coopervision Contact lens parameters

Coopervision Monthly contact lenses 2/2

| Contact Lens Name | Contact Lens Specifications | Contact Lens powers | Astigmatism powers and axis available |
|-----------------------------|---|--|---|
| Biofinity | Base Curve: 8.6 Diameter: 14.0 Dk/t @-3.00D 160 | +0.25D to +6.00D (0.25D steps) +6.00D to +8.00D (0.50D steps) plano to -6.00D (0.25D steps) -6.00D to -12.00D (0.50D steps) | |
| Biofinity Toric | Base Curve: 8.7 Diameter: 14.5 Dk/t @-3.00D 116 | +0.25D to +6.00D (0.25D steps) +6.00D to +8.00D (0.50D steps) plano to -6.00D (0.25D steps) -6.00D to -10.00D (0.50D steps) | Cyl Powers: +0.75, +1.25, +1.75, +2.25 Axis: Full Circle in 10 deg steps |
| Biofinity Multifocal | Base Curve: 8.6 Diameter: 14.0 Dk/t @-3.00D 142 Centre Distance, Centre Near | +0.25D to +6.00D (0.25D steps) plano to -6.50D (0.25D steps) -6.00D to -8.00D (0.50D steps) ADD powers +1.00, +1.50,+2.00,+2.50 | |
| Avaira | BC: 8.5 (-) & 8.4 (+) Diameter: 14.2 Dk/t @-3.00D 125 | Base Curve 8.4: +0.25D to +6.00D(0.25D steps) +6.00D to +8.00D (0.50D steps) Base Curve 8.5: -0.25D to -6.00D (0.25D steps) -6.00D to -12.00D (0.50D steps) | |
| Avaira Toric | Base Curve: 8.5 Diameter: 14.5 Dk/t @-3.00D 91 | +0.25D to +6.00D (0.25D steps) plano to -6.00D (0.25D steps) -6.50D to -10.00D (0.50D steps) | Cyl : -0.75, -1.25, -1.75, -2.25 Axis: Full Circle (10 deg steps) |



<http://optometryportal.com/>
Follow on Twitter @optometryportal
Download the 'Optometry portal' Android App